



Nederlands Meetinstituut

Test certificate

Number TC2317 Revision 2
Project number 10030701
Page 1 of 4

Issued by NMI IJkwezen B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
The Netherlands

Notified Body Number 122

In accordance with Paragraph 8.1 of the European Standard on Metrological aspects of non-automatic weighing instrument EN 45501:1992 and by application of the OIML International Recommendation R 60 (Edition 1991). The applied error fraction p meant in paragraph 3.5.4. of the standard is 0.7.

Applicant Celtron Technologies Inc
No. 1 Lane 86 Section 1
Pa Lien Road, HSI TZU
Taipei Hsien
TAIWAN

In respect of The model of a **load cell** with strain gauges, tested as part of a weighing instrument:
manufacturer : Celtron
type : LPS

Characteristics

Maximum Capacity (E_{max})	6, 10, 15, 20 and 35 kg
Accuracy Class	C
Maximum number of load cell intervals (n)	3000
Minimum load cell verification interval (V_{min})	$E_{max} / 5000$

In the description TC2317 revision 1 further essential characteristics are described.

Description and Documentation The load cell is described in the description number TC2317 revision 1 and documented in the documentation folder number TC2317-2, both appertaining to this test certificate number TC2317.

Remarks Summary of tests involved: see Appendix number TC2317 Revision 1.
This test certificate replaces test certificate with number TC2317 Revision 1 except for the documentation folder number TC2317-2.

Dordrecht, 19 September 1994

NMI IJkwezen B.V.

M. Charité
Director

Nederlands Meetinstituut
Hugo de Grootplein 1
3314 EG Dordrecht (NL)
Telephone (078) 33 23 32
Telefax (078) 33 23 05

Nederlands Meetinstituut N.V. (Registered at the
Chamber of Commerce Delft number 28701)

Subsidiary companies:
NMI Van Swinden Laboratorium B.V. (28703)
NMI IJkwezen B.V. (28700)
NMI Certis B.V. (33418)
NMI Test- en Afdeling (TAC) B.V. (28702)

This certificate is issued under the provision that Nederlands Meetinstituut N.V. nor its subsidiary companies accept any liability.

Reproduction of the complete certificate is allowed.
Parts of the certificate may only be reproduced after written permission.